

LIFEPAK 15 ECG Transmissions

During time-critical pre-hospital care, Avel eCare EMS stands ready to expand your EMS team to include ECG interpretation in coordination of care while you remain focused on your patient.

Simple Workflow Changes for Crews to Know

Please educate the EMS team of the process for an ECG interpretation by an Avel eCare EMS Physician.

1. EMS Provider must activate Avel eCare EMS prior to transmitting ECG.
2. Patient identifiers are not required to be entered into the LIFEPAK 15 prior to transmission.
3. EMS Provider to request ECG read and communicate transmission status via phone activation or during live video encounter.
4. Avel eCare EMS Physician to interpret ECG and provide Medical Control support in the treatment plan.
5. EMS Provider and Avel eCare EMS collaborate to confirm if Avel eCare EMS should transmit the ECG to the receiving facility.

Preparing Your Agency's LIFEPAK 15

For Agencies with a new LIFEPAK 15 where Avel eCare EMS is programmed:

1. Create and/or activate Avel eCare EMS as a receiving target in your Agency's LIFENET Account. If support is needed, contact Stryker Data Solutions Support by calling 844-999-2787.
2. Ensure the ECG cable is plugged into the front of the LIFEPAK.
3. Power on the device and select 12-lead. The device will notify you the leads are off, select the 12-lead button a second time and a blank 12-lead will print.
4. Once printing starts, select Transmit > AVEL (report type 12-lead) > Send.

For Agencies with a pre-existing LIFEPAK 15 where Avel eCare EMS is not programmed:

1. Create and/or activate Avel eCare EMS as a receiving target in your Agency's LIFENET Account. If support is needed, contact Stryker Data Solutions Support by calling 844-999-2787.
2. Turn on the LIFEPAK in Set-up mode by selecting Options/Events/On. The 4-digit code is 0000 unless changed by your local team.
3. Once in Set-up mode, select Transmission > Sites > and the 1st available site.
 - o Add in AVEL (exactly) > select EMD > Default output is Direct Connect
4. Select previous page and power off the device.
5. Ensure the ECG cable is plugged into the front of the device.
6. Power on the device and select 12-lead. The device will notify you the leads are off, select the 12-lead button a second time and a blank 12-lead will print.
7. Once printing starts, select Transmit > AVEL (report type 12-lead) > Send.